W01914201156E

检 测 报 告

Test Report

产 品 名 称 ： **Digital Load Cell**

Name of Sample

型 号 规 格 ： **ZSW**

Type

委 托 单 位 ： **Keli Sensing Technology (Ningbo) Co.,Ltd.**

Applicant

检 测 类 别 ： **Entrusted Test**

Test Purpose

上海市质量监督检验技术研究院

Shanghai Institute of Quality Inspection and Technical Research

Shanghai Institute of Quality Inspection and Technical Research
Testing Report

No.: W01914201156E

Name of Sample	Digital Load Cell	Test Purpose	Entrusted Test
Type Grade	ZSW Good samples	Trade Mark	
Applicant	Keli Sensing Technology (Ningbo) Co.,Ltd.		
Tested Company	Keli Sensing Technology (Ningbo) Co.,Ltd.		
Producer	Keli Sensing Technology (Ningbo) Co.,Ltd.		
Number of Client	xxxxxxx	Entrusting Sampling Date	2019.10.22
Reception Date	2019.10.22	Sampling Spot	/
Sample Quantity	1	Sum of Sample	/
Date of Production	/	Original Number	/
Situation of Sample	Good (send by client)		
Testing Place	No.900 Jiangyue Road, Shanghai 201114, P.R.CHINA		
Test Standard/ Judgment Rules	Per client's requirements		
Date of Testing	2019.10.22~2019.10.23		
Conclusion	The tested results are accord with the test standards and / or judge basis for the tested items. Please see the results of this report. Issued Date： 2019.10.23		
Client's Message	Add.	No：199 ChangXing Road,Jiangbei District,Ningbo,China	
	Zip Code	/	Tel. /
Note	1. The test report only offers the conclusion for the tested items required in the related test standards and / or judge basis. 2. The tests are not accredited, they shall only be used for technical research, education or internal quality control purpose.		

Approved By: Lubin Chief Lu Bin Checked By: Shen haizhou Tested by: Xu Yi

Collection of The Test Results					
No.	Test items	Technical Requirements	Test Results	Judgements by Single item	Remarks
1	Lighting stroke test	Applying 80kA, 8/20us surge one time ; Check there is no obvious black or burning phenomenon of the sample during the test Check there is no visible damage of sample after test Check the function is normal	Pass	Accord	Details see annex

Annex

1.Lighting stroke test

Sample Description: Digital Load Cell

Sample Quantity: 1

Sample No. : 1#

Test Condition：

Test methods:

Applying current wave test according to requirement.

Test current: 80KA

Test wave: （8/20us）


Test time: one times

Requirement:

No obvious black or burning phenomenon of the sample during the test.
No visible damage of sample after test.
The communication with instrument and the function should be normal.

Test Result:

Sample No.	Test current	Test result	Conclusion
1#	80kA	No obvious black or burning phenomenon of the sample during the test; No visible damage of sample after test The communication with instrument and the function are normal.	Pass



Test wave picture of the ZSW in 8/20us 80kA

Sample picture



The below is blank